

Despite these caveats, Zhang and Hei's point is well made and highlights the work that is still required to successfully advance cell therapy for ARDS, especially in the context of extracorporeal organ support. We hope that our study illustrates the usefulness of clinically relevant, high-fidelity animal models in advancing these efforts. ■

Author disclosures are available with the text of this letter at www.atsjournals.org.

Jonathan E. Millar, M.D.*
University of Edinburgh
Edinburgh, United Kingdom
and
University of Queensland
Brisbane, Queensland, Australia

Jacky Y. Suen, Ph.D.
University of Queensland
Brisbane, Queensland, Australia

Daniel F. McAuley, M.D.
Queen's University Belfast
Belfast, United Kingdom

John F. Fraser, M.D., Ph.D.
University of Queensland
Brisbane, Queensland, Australia

On behalf of all the authors

ORCID ID: 0000-0002-4853-9377 (J.E.M.).

*Corresponding author (e-mail: j.millar@doctors.org.uk).

References

1. Millar JE, Bartnikowski N, Passmore MR, Obonyo NG, Malfertheiner MV, von Bahr V, *et al*. Combined mesenchymal stromal cell therapy and extracorporeal membrane oxygenation in acute respiratory distress syndrome: a randomized controlled trial in sheep. *Am J Respir Crit Care Med* 2020;202:383–392.
2. Abraham A, Krasnodembskaya A. Mesenchymal stem cell-derived extracellular vesicles for the treatment of acute respiratory distress syndrome. *Stem Cells Transl Med* 2020;9:28–38.
3. Xu B, Chen SS, Liu MZ, Gan CX, Li JQ, Guo GH. Stem cell derived exosomes-based therapy for acute lung injury and acute respiratory distress syndrome: a novel therapeutic strategy. *Life Sci* 2020;254:117766.
4. Emukah C, Dittmar E, Naqvi R, Martinez J, Corral A, Moreira A, *et al*. Mesenchymal stromal cell conditioned media for lung disease: a systematic review and meta-analysis of preclinical studies. *Respir Res* 2019;20:239.
5. Kraitchman DL, Tatsumi M, Gilson WD, Ishimori T, Kedziorek D, Walczak P, *et al*. Dynamic imaging of allogeneic mesenchymal stem cells trafficking to myocardial infarction. *Circulation* 2005;112:1451–1461.
6. Sengupta V, Sengupta S, Lazo A, Woods P, Nolan A, Bremer N. Exosomes derived from bone marrow mesenchymal stem cells as treatment for severe COVID-19. *Stem Cells Dev* 2020;29:747–754.
7. Islam MN, Das SR, Emin MT, Wei M, Sun L, Westphalen K, *et al*. Mitochondrial transfer from bone-marrow-derived stromal cells to pulmonary alveoli protects against acute lung injury. *Nat Med* 2012;18:759–765.
8. Hayes M, Curley GF, Masterson C, Devaney J, O'Toole D, Laffey JG. Mesenchymal stromal cells are more effective than the MSC secretome in diminishing injury and enhancing recovery following ventilator-induced lung injury. *Intensive Care Med Exp* 2015;3:29.
9. Silva JD, de Castro LL, Braga CL, Oliveira GP, Trivelin SA, Barbosa-Junior CM, *et al*. Mesenchymal stromal cells are more effective than their extracellular vesicles at reducing lung injury regardless of acute respiratory distress syndrome etiology. *Stem Cells Int* 2019;2019:8262849.
10. Silachev DN, Goryunov KV, Shpilyuk MA, Beznoschenko OS, Morozova NY, Kraevaya EE, *et al*. Effect of MSCs and MSC-derived extracellular vesicles on human blood coagulation. *Cells* 2019;8:258.
11. Fiedler T, Rabe M, Mundkowski RG, Oehmcke-Hecht S, Peters K. Adipose-derived mesenchymal stem cells release microvesicles with procoagulant activity. *Int J Biochem Cell Biol* 2018;100:49–53.

Copyright © 2020 by the American Thoracic Society



Considerations for an Optimal Electrical Activity of the Diaphragm Threshold for Automated Detection of Ineffective Efforts

To the Editor:

We have read with great interest the research letter authored by Jonkman and colleagues (1), and we agreed with the notion that suboptimal filtering of the electrical activity of the diaphragm (EAdi) signal together with a low threshold ($>1 \mu\text{V}$), could lead to incorrect interpretation of patient–ventilator interactions when detected by automated software.

In our reported validation investigation of Better Care software (2), the algorithm performance was made against five different experts' opinions using 1,024 tracings of airway flow and airway pressure waveform from 16 different patients, with a reported sensitivity of 91.5% and specificity of 91.7%. Subsequently, as an additional confirmation, we used EAdi tracings with a threshold $>1 \mu\text{V}$ in eight mechanically ventilated patients, obtaining a sensitivity of 65.2% and a specificity of 99.3%. This value was selected on an *a priori* basis, considering a midpoint between $0.1 \mu\text{V}$ and $2 \mu\text{V}$ and was intended to avoid inspiratory assistance during expiration in those cases when the EAdi peak is $<1.5 \mu\text{V}$ and the cycling-off is at a 40% threshold from EAdi peak, instead of the usual 70% (3).

The drop in sensitivity of Better Care algorithm when EAdi was used could be due to, as the authors speculate, a mistaken overestimation of ineffective efforts by EAdi, leading to an increase in false-negative results in the Better Care algorithm. We have seriously considered this possibility in those tracings validated against EAdi, and we have reanalyzed tracings from that previously published cohort, searching for the best cutoff value of EAdi signal with the best performance. The new findings show that the best cutoff value of EAdi is $2.3 \mu\text{V}$, with a sensitivity of 89.2%, a specificity of 96%, a positive predictive value of 72.5%, and a negative predictive value of 98.7%.

†This article is open access and distributed under the terms of the Creative Commons Attribution Non-Commercial No Derivatives License 4.0 (<http://creativecommons.org/licenses/by-nc-nd/4.0/>). For commercial usage and reprints, please contact Diane Gern (dgem@thoracic.org).

Originally Published in Press as DOI: 10.1164/rccm.202007-2960LE on September 16, 2020

Overall, it seems that increasing the threshold of EAdi would decrease the false-negative rate, improving the sensitivity of any given automated detection software and keeping a good specificity. We believe that, according to our reassessed results, an EAdi $>2 \mu\text{V}$ could be suitable for this purpose. In addition, as Jonkman and colleagues mentioned, the removal of cardiac electrical activity is technically challenging, particularly when the signal:noise ratio of the crural diaphragm electromyography signal is low. In this scenario, we hypothesized that the automatic detection of true ineffective efforts from EAdi will be improved by using a personalized adaptive threshold for each patient considering the signal:noise ratio of the diaphragm electromyography signal. Interestingly, nonlinear methods less sensitive to ECG interference based on sample entropy algorithms (4) could be used to reduce the delay on the neural onset when an ECG peak matches at the beginning of the breath. ■

Author disclosures are available with the text of this letter at www.atsjournals.org.

José Aquino-Esperanza, M.D.*
Hospital Universitari Parc Taulí
Sabadell, Spain

Centro de Investigación Biomédica en Red Enfermedades Respiratorias
(CIBERES)
Madrid, Spain

and
Universitat de Barcelona
Barcelona, Spain

Leonardo Sarlabous, Ph.D.
Hospital Universitari Parc Taulí
Sabadell, Spain

Rudys Magrans, Ph.D.
Jaume Montanya, M.Sc.
Better Care SL
Sabadell, Spain

Umberto Lucangelo, M.D., Ph.D.
Trieste University
Trieste, Italy

Lluís Blanch, M.D., Ph.D.
Hospital Universitari Parc Taulí
Sabadell, Spain

and
Centro de Investigación Biomédica en Red Enfermedades Respiratorias
(CIBERES)
Madrid, Spain

ORCID ID: 0000-0002-2626-7593 (J.A.-E.).

*Corresponding author (e-mail: jaquino@tauli.cat).

References

1. Jonkman AH, Roesthuis LH, de Boer EC, de Vries HJ, Girbes ARJ, van der Hoeven JG, *et al*. Inadequate assessment of patient-ventilator interaction due to suboptimal diaphragm electrical activity signal filtering. *Am J Respir Crit Care Med* 2020;202:141–144.
2. Blanch L, Sales B, Montanya J, Lucangelo U, Garcia-Esquirol O, Villagra A, *et al*. Validation of the Better Care® system to detect ineffective efforts during expiration in mechanically ventilated patients: a pilot study. *Intensive Care Med* 2012;38:772–780.
3. Suarez-Sipmann F, Pérez Márquez M, González Arenas P. New modes of ventilation: NAVA [in Spanish]. *Med Intensiva* 2008;32:398–403.
4. Sarlabous L, Estrada L, Cerezo-Hernández A, Leest SVD, Torres A, Jan R, *et al*. Electromyography-based respiratory onset detection in COPD patients on non-invasive mechanical ventilation. *Entropy (Basel)* 2019; 21:258.

Copyright © 2020 by the American Thoracic Society



Reply to Aquino-Esperanza *et al*.



From the Authors:

We greatly appreciate the interest of Aquino-Esperanza and colleagues in our research letter (1) regarding the influence of suboptimal filtering of the electrical activity of the diaphragm (EAdi) signal on the detection of patient-ventilator asynchronies. In that letter, we raised the concern that cardiac activity-related artifacts in the EAdi signal may be mistakenly detected as ineffective efforts when the EAdi threshold is too low. Based on this work, Aquino-Esperanza and colleagues have thoughtfully reanalyzed the performance of their Better Care algorithm (2) to find an appropriate EAdi threshold for the automatic detection of ineffective efforts. They conclude that increasing the EAdi threshold from $1 \mu\text{V}$ to $2.3 \mu\text{V}$ improved the sensitivity of their algorithm and maintained adequate specificity. We appreciate this careful reanalysis and agree with the authors that a threshold $>2 \mu\text{V}$ is reasonable. It should be noted that in our work (1), EAdi artifacts were mostly $<4 \mu\text{V}$, but we agree that a threshold of $4 \mu\text{V}$ would be clinically disproportionate and increase the false-negative rate.

We also agree with the authors that a personalized adaptive EAdi threshold may improve the performance of automatic detection of true ineffective efforts. We considered testing this with our dataset; however, the incidence of true ineffective efforts was too low (1). In contrast, because the processing of these EAdi artifacts is a technical issue, it might be rather impossible to distinguish between artifacts and true ineffective efforts based on a certain EAdi threshold solely. As part of our earlier work, we aimed to quantify waveform characteristics of the cardiac activity-related artifacts (e.g., slope of the inspiratory EAdi increase, timing, and amplitude) and predict the occurrence of these artifacts based on patient characteristics. For instance, we hypothesized that cardiac activity-related peaks had steeper increases (“sharp waves”, possibly consistent with fast cardiac depolarization); however, slopes of the artificial and true peaks were similar on average, and artificial peaks with both lower and higher slopes compared with true EAdi peaks were found within patients. Furthermore, factors such as the presence of ventricular hypertrophy were not related to the occurrence of these artifacts. We did not include these findings in our research letter, as this is clinically not very helpful at this time.

Importantly, the main challenge with developing a (personalized) EAdi threshold using signal characteristics is the

†This article is open access and distributed under the terms of the Creative Commons Attribution Non-Commercial No Derivatives License 4.0 (<http://creativecommons.org/licenses/by-nc-nd/4.0/>). For commercial usage and reprints, please contact Diane Gern (dgem@thoracic.org).

Originally Published in Press as DOI: 10.1164/rccm.202008-3052LE on September 16, 2020